



DEKS

Rigshospitalet - Glostrup
Valdemar Hansens Vej 1-23
Opgang 8,1.
2600 Glostrup
Denmark
☎ +45 3863 4400
www.deks.dk

Test period:
18.11.2019 – 02.12.2019
Last reporting day:
02.12.2019

Program coordinator:
Sanne Schou
☎ +45 38634405
sanne.schou@deks.dk
Gitte M. Henriksen
☎ +45 38634401
gitte.henriksen@deks.dk

This report contains:
• This cover letter
• Individual results
• Guide to interpretation

Approved by GMH
09.12.2019

Next reporting period

10.02.2020



copyright © DEKS

Homocysteine, HCY 3034DK EQA rapport 5/ 2019

Participants

There are 82 laboratories registered for this scheme. Most of the participants are from the Scandinavian countries and 11 laboratories are from other countries. DEKS has received 80 results for this report.

Samples

Sample no. 09_2019 "Glp" is a pool of EDTA plasma. This pool has been sent out twice before.

Sample no. 10_2019 "Heg" is a pool of EDTA plasma. This pool has been sent out once before.

Statistic

In the graphical part you will find your own results and deviation (dev.) from the mean (M).

Target value

The target value is calculated as the mean of all.

Acceptance

The acceptance interval is $\pm 17\%$.

Results and comments

In this round no outliers was found. Outliers are defined as $>SD*3,6$ from the overall mean.

Kind regards

Gitte Henriksen og Sanne Schou

1) The acceptance interval is calculated based on biological variation from Carmen Ricos TE%. <https://www.westgard.com/biodatabase1.htm>

