



## DEKS

Rigshospitalet - Glostrup  
Valdemar Hansens Vej 1-23  
Opgang 8, 1.  
2600 Glostrup  
Denmark  
☎ +45 3863 4400  
www.deks.dk

Test period:  
07.02.2020 – 24.02.2020  
Last reporting day:  
24.02.2020

Program coordinator:  
Sanne Schou  
☎ +45 38634405  
sanne.schou@deks.dk  
Gitte M. Henriksen  
☎ +45 38634401  
gitte.henriksen@deks.dk

This report contains:  
• This cover letter  
• Individual results

Approved by GMH  
02.03.2020

## Next reporting period

27.04.2020



PT Reg.no. 591

copyright © DEKS

# Homocysteine, HCY 3034DK EQA rapport 1/ 2020

## Participants

There are 88 laboratories registered for this scheme. Most of the participants are from the Scandinavian countries and 11 laboratories are from other countries.

DEKS has received 89 results for this report.

## Samples

Sample no. 01\_2020 "Goq" is a pool of EDTA plasma. This pool has been sent out twice before.

Sample no. 02\_2020 "Amr" is a pool of EDTA plasma. This pool has been not been sent out before.

## Statistic

In the graphical part you will find your own results and deviation (dev.) from the mean (M).

### Target value

The target value is calculated as the mean of all.

### Acceptance

The acceptance interval is  $\pm 17\%$ <sup>1</sup>.

## Results and comments

In this round no outliers were found. Outliers are defined as  $>SD*3,6$  from the overall mean.

## Kind regards

Gitte Henriksen og Sanne Schou

1) The acceptance interval is calculated based on biological variation from Carmen Ricos TE%. <https://www.westgard.com/biodatabase1.htm>

