



DEKS

Rigshospitalet - Glostrup
Valdemar Hansens Vej 1-23
Opgang 8, 1.
2600 Glostrup
Denmark
☎ +45 3863 4400
www.deks.dk

Test period:
27.05.2020 – 11.05.2020
Last reporting day:
11.05.2020

Program coordinator:
Sanne Schou
☎ +45 38634405
sanne.schou@deks.dk
Gitte M. Henriksen
☎ +45 38634401
gitte.henriksen@deks.dk

This report contains:
• This cover letter
• Individual results

Approved by GMH
19.05.2020

Next reporting period

17.08.2020



copyright © DEKS

Homocysteine, HCY 3034DK EQA rapport 2/ 2020

Participants

There are 88 laboratories registered for this scheme. Most of the participants are from the Scandinavian countries and 18 laboratories are from other countries.

DEKS has received 86 results for this report.

Samples

Sample no. 03_2020 "Gmn" is a pool of EDTA plasma. This pool has been sent out three times before.

Sample no. 04_2020 "Bch" is a pool of EDTA plasma. This pool has been not been sent out before.

Statistic

In the graphical part you will find your own results and deviation (dev.) from the mean (M).

Target value

The target value is calculated as the mean of all.

Acceptance

The acceptance interval is $\pm 17\%$ ¹.

Results and comments

In this round two outliers were found. Outliers are defined as $>SD*3,6$ from the overall mean.

Kind regards

Gitte Henriksen og Sanne Schou

1) The acceptance interval is calculated based on biological variation from Carmen Ricos TE%. <https://www.westgard.com/biodatabase1.htm>

