

Homocysteine, HCY 3034DK EQA rapport 4/ 2021

DEKS

Rigshospitalet - Glostrup
Valdemar Hansens Vej 1-23
Opgang 8,1.
2600 Glostrup
Denmark
☎ +45 3863 4400
www.deks.dk

Test period:
11.10.2021 – 02.11.2021
Last reporting day:
02.11.2021

Scheme coordinator:
Sanne Schou
☎ +45 38634405
sanne.schou@deks.dk
Gitte M. Henriksen
☎ +45 38634401
gitte.henriksen@deks.dk

This report contains:

- This letter
- Summary report
- Individual results

Approved by GMH
05.11.2021

Next reporting period starts

29.11.2021



PI Reg.no. 591

copyright © DEKS

Participants

There are 86 laboratories registered for this scheme. Most of the participants are from the Scandinavian countries and 14 laboratories are from other countries.

DEKS has received 78 results for this report.

Samples

Sample no. 07_2021 "Bch" is a pool of EDTA plasma. This pool has been sent out twice before.

Sample no. 08_2021 "Ept" is a pool of EDTA plasma. This pool has not been sent out before.

Statistic

In the graphical part you will find your own results and deviation (dev.) from the mean (M).

Target value

The target value is calculated as the mean of all.

Acceptance

The acceptance interval is $\pm 17\%$ ¹.

Results and comments

In this round no outliers were found. Outliers are defined as $>SD*3,6$ from the overall mean.

Kind regards

Gitte Henriksen og Sanne Schou

1) The acceptance interval is calculated based on biological variation from Carmen Ricos TE%. <https://www.westgard.com/biodatabase1.htm>