

Lupus Anticoagulant

Ratio Screening/Confirmation

No. of Responders  
486

Sample No 16.49  
Sample Details Plasma weakly positive for Lupus Anticoagulant  
Units: Ratio  
Prior Use: None

Stability 14-February-2017

Homogeneity 0.0 %

Homogeneity Parameter

Lupus ratio

Ratio Screening/Confirmation

	n	assigned value	CV (%)	range	panel 1	z-score	panel 2	z-score	panel 3	z-score
APTT	72	1.25	13.7	0.10 - 7.00						
dAPTT	6	1.38		1.13 - 1.52						
dPT	5	1.16		1.08 - 1.56						
dRVVT	417	1.42	8.4	0.86 - 2.12			1.41	-0.08		
SCT	78	1.83	14.7	1.16 - 5.96					1.76	-0.25

Assays

