

CN01 - Clobazam, Norclobazam

Analyte	Analyst	Method	Result	Units	Z score (** z' score)	Assigned Value	Ux AV	SDPA	Exp.SDPA	No of results	Median	Mean	Robust SD	SD
CN1A - Clobazam	Lab Result	LCMS & MS/MS	573.00	µg/L	0.37	552.00	10.65	56.737	N/A	19	552.00	543.75	47.456	51.381
CN1B - Clobazam	Lab Result	LCMS & MS/MS	175.00	µg/L	-0.82	192.00	5.16	20.652	N/A	18	188.60	188.47	20.169	20.420
CN1A - Norclobazam	Lab Result	LCMS & MS/MS	1,971.00	µg/L	0.74	1,857.70	24.86	153.498	N/A	15	1,857.00	1,827.87	135.561	172.013
CN1B - Norclobazam	Lab Result	LCMS & MS/MS	244.00	µg/L	-0.15	249.43	4.28	37.410	N/A	14	249.00	235.95	18.538	43.355

PS23 - Zuclophenithoxol

Analyte	Analyst	Method	Result	Units	Z score (** z' score)	Assigned Value	Ux AV	SDPA	Exp.SDPA	No of results	Median	Mean	Robust SD	SD
Zuclophenithoxol	Lab Result	LCMS & MS/MS	4.08	µg/L	-0.71	4.81	0.17	1.025	N/A	29	4.81	4.75	0.601	0.864

** Please note, participant performance for this analyte has been assessed using a z' score, rather than a z score, in order to account for the measurement uncertainty of the assigned value which is not negligible when compared to the SDPA.