

AE5 - Anti-epileptic drugs 5

Analyte	Analyst	Method	Result	Units	Z Score	Assigned Value	Ux AV	SDPA	Exp.SDPA	Number of results	Median	Mean	Robust SD	SD	Your Reference
Brivaracetam	Lab Result	LCMS & MS/MS	3.13	mg/L	0.97*	2.89	0.07	0.24	0.25	17	2.89	2.89	0.237	0.224	LCMS & MS/MS

* Please note, participant performance for this analyte has been assessed using a z' score, rather than a z score, in order to account for the measurement uncertainty of the assigned value which is not negligible when compared to the SDPA.

CN01 - Clobazam, Norclobazam

Analyte	Analyst	Method	Result	Units	Z Score	Assigned Value	Ux AV	SDPA	Exp.SDPA	Number of results	Median	Mean	Robust SD	SD	Your Reference
CN1A - Clobazam	Lab Result	LCMS & MS/MS	232.00	µg/L	-1.26	265.50	5.25	26.57	N/A	30	265.50	264.85	22.987	27.039	LCMS & MS/MS
CN1B - Clobazam	Lab Result	LCMS & MS/MS	627.00	µg/L	-1.40	744.00	17.59	83.28	N/A	29	744.00	721.64	75.781	101.744	LCMS & MS/MS
CN1A - Norclobazam	Lab Result	LCMS & MS/MS	425.00	µg/L	-0.82	464.45	7.68	48.38	N/A	25	464.45	470.41	30.105	70.934	LCMS & MS/MS
CN1B - Norclobazam	Lab Result	LCMS & MS/MS	3317.00	µg/L	-0.52	3499.17	104.82	353.11	N/A	24	3499.17	3510.68	402.150	387.852	LCMS & MS/MS

PS29 - Trazodone

Analyte	Analyst	Method	Result	Units	Z Score	Assigned Value	Ux AV	SDPA	Exp.SDPA	Number of results	Median	Mean	Robust SD	SD	Your Reference
Trazodone	Lab Result	LCMS & MS/MS	620.00	µg/L	1.57	529.00	13.91	57.84	N/A	28	529.00	547.17	57.837	61.496	LCMS & MS/MS