

CN01 - Clobazam, Norclobazam

Analyte	Analyst	Method	Result	Units	Z score (** z' score)	Assigned Value	Ux AV	SDPA	Exp.SDPA	No of results	Median	Mean	Robust SD	SD
CN1A - Clobazam	Lab Result	LCMS & MS/MS	233.00	µg/L	-0.52	242.00	3.64	17.247	N/A	22	244.00	246.87	20.762	26.264
CN1B - Clobazam	Lab Result	LCMS & MS/MS	<10.00	µg/L		Zero Spike	N/A	N/A	N/A	20	N/A	N/A	N/A	N/A
CN1A - Norclobazam	Lab Result	LCMS & MS/MS	464.00	µg/L	0.67	447.13	6.17	25.174	N/A	16	462.77	454.21	34.851	36.725
CN1B - Norclobazam	Lab Result	LCMS & MS/MS	<10.00	µg/L		Zero Spike	N/A	N/A	N/A	15	N/A	N/A	N/A	N/A

PS07 - Fluphenazine

Analyte	Analyst	Method	Result	Units	Z score (** z' score)	Assigned Value	Ux AV	SDPA	Exp.SDPA	No of results	Median	Mean	Robust SD	SD
Fluphenazine	Lab Result	LCMS & MS/MS	1.22	µg/L	-0.64**	1.36	0.07	0.208	0.219	13	1.35	1.37	0.208	0.204

** Please note, participant performance for this analyte has been assessed using a z' score, rather than a z score, in order to account for the measurement uncertainty of the assigned value which is not negligible when compared to the SDPA.